



EURADOS IC2012n

Irradiations at NPL

Graeme C. Taylor
Neutron Metrology Group, NPL

June 4th 2013




- **NPL Irradiation Facility**
- **Dosimeter Storage**
- **Irradiation Details**
- **Exposure Validation**
- **Summary of Uncertainties**



Chadwick Building




Low Scatter Facility



Dosimeter Storage



Dosimeters bagged as unpacked and placed in storage units



Irradiation Details



Source emission rates measured in NPL Manganese Bath
Source anisotropy measured in same facility as exposures



Dose (mSv)	Angle	Source Type	Source ID	Emission rate* (s ⁻¹)	% of ²⁵⁰ Cf	Exposure time (hh:mm)
0.3	0°	²⁵² Cf (bare)	5000NC	3.4x10 ⁷	13%	00:26
3	0°	²⁵² Cf (bare)	1254NU	2.9x10 ⁸	negligible	00:30
15	0°	²⁵² Cf (bare)	1254NU	"	"	02:27
2	45°	²⁵² Cf (bare)	1254NU	"	"	00:20
3	0°	²⁵² Cf (moderated)	1254NU	"	"	02:06

*On 31/10/2012



Reference EPD-N2s mounted on top of phantoms during exposures

7200153_P26_3mSv.txt - Notepad

File Edit Format View Help

Record,	DateTime,	Interval,	EpdClock Sec,	HpG uSv,	HpN uSv
000,	24/10/2012 20:06:11,	00:00:30,	65481241,	109,	1466
001,	24/10/2012 20:05:41,	00:00:30,	65481211,	108,	1465
002,	24/10/2012 20:05:11,	00:00:30,	65481181,	106,	1465
003,	24/10/2012 20:04:41,	00:00:30,	65481151,	104,	1434
004,	24/10/2012 20:04:11,	00:00:30,	65481121,	103,	1434
005,	24/10/2012 20:03:41,	00:00:30,	65481091,	101,	1424
006,	24/10/2012 20:03:11,	00:00:30,	65481061,	99,	1404
007,	24/10/2012 20:02:41,	00:00:30,	65481031,	97,	1384
008,	24/10/2012 20:02:11,	00:00:30,	65481001,	95,	1364
009,	24/10/2012 20:01:41,	00:00:30,	65480971,	93,	1344
010,	24/10/2012 20:01:11,	00:00:30,	65480941,	91,	1324
011,	24/10/2012 20:00:41,	00:00:30,	65480911,	89,	1304
012,	24/10/2012 20:00:11,	00:00:30,	65480881,	87,	1284
013,	24/10/2012 19:59:41,	00:00:30,	65480851,	85,	1264
014,	24/10/2012 19:59:11,	00:00:30,	65480821,	83,	1244
015,	24/10/2012 19:58:41,	00:00:30,	65480791,	81,	1224
016,	24/10/2012 19:58:11,	00:00:30,	65480761,	79,	1204
017,	24/10/2012 19:57:41,	00:00:30,	65480731,	77,	1184
018,	24/10/2012 19:57:11,	00:00:30,	65480701,	75,	1164
019,	24/10/2012 19:56:41,	00:00:30,	65480671,	73,	1144
020,	24/10/2012 19:56:11,	00:00:30,	65480641,	71,	1124
021,	24/10/2012 19:55:41,	00:00:30,	65480611,	69,	1104
022,	24/10/2012 19:55:11,	00:00:30,	65480581,	67,	1084
023,	24/10/2012 19:54:41,	00:00:30,	65480551,	65,	1064
024,	24/10/2012 19:54:11,	00:00:30,	65480521,	63,	1044
025,	24/10/2012 19:53:41,	00:00:30,	65480491,	61,	1024
026,	24/10/2012 19:53:11,	00:00:30,	65480461,	59,	1004
027,	24/10/2012 19:52:41,	00:00:30,	65480431,	57,	984
028,	24/10/2012 19:52:11,	00:00:30,	65480401,	55,	964
029,	24/10/2012 19:51:41,	00:00:30,	65480371,	53,	944
030,	24/10/2012 19:51:11,	00:00:30,	65480341,	51,	924
031,	24/10/2012 19:50:41,	00:00:30,	65480311,	49,	904
032,	24/10/2012 19:50:11,	00:00:30,	65480281,	47,	884
033,	24/10/2012 19:49:41,	00:00:30,	65480251,	45,	864
034,	24/10/2012 19:49:11,	00:00:30,	65480221,	43,	844
035,	24/10/2012 19:48:41,	00:00:30,	65480191,	41,	824

EasyEPD2 EPD-N2 ID: 07200154 Mk2.50 Software Version 5

File View SetUp Window Help

Dose and Alarms

Wearer Name: ID: FFFFFFFFFF

Dose and Rates

	Dose	Total	Rate	Peak	Peak Rate Time
	uSv	uSv	uSv/h	uSv/h	
HpG+HpN	2106.14				
HpG	360.31	360.31	1	200	13/11/2012 12:11:35
HpN	1745.83	1745.83	300	1100	13/11/2012 12:24:32

Counts Since 13/11/2012 11:39:29

HG 45804 SG 10116 FN 138 AN 367

Dose Quality

Dose Alarm Thresholds

	uSv
HpG (1)	100000.00
HpG (2)	
HpN	100000.00
HpG+HpN	100000.00

Rate Alarm Thresholds

Off	On
uSv/h	uSv/h
9000000	10000000
9000000	10000000
40000000	50000000

EPD Issue De-Issue Off Clear All Counts Peak Rates Dose+Q'ty Clear Total Close

Summary of Uncertainties



Table 2: Percentage standard uncertainties associated with the determination of the personal dose equivalent at the reference distance.

Uncertainty component	Irradiation				
	²⁵² Cf 0° 0.3 mSv	²⁵² Cf, 0° 3 mSv	²⁵² Cf 0° 15 mSv	²⁵² Cf(D ₂ O) 0° 3 mSv	²⁵² Cf 45° 2 mSv
Type B (non-random)					
Reference irradiation distance*	± 0.53%	± 0.53%	± 0.53%	± 0.53%	± 0.53%
Source emission rate (MnSO ₄ bath) (includes component for half-life)	± 0.60%	± 0.40%	± 0.40%	± 0.40%	± 0.40%
Source anisotropy correction	± 0.50%	± 0.50%	± 0.50%	± 0.0%	± 0.50%
Timing	± 0.26%	± 0.22%	± 0.04%	± 0.05%	± 0.33%
Scatter	± 1.0%	± 1.0%	± 1.0%	± 1.0%	± 1.0%
H _p (10,θ) conversion coefficient	± 1.0%	± 1.0%	± 1.0%	± 4.0%	± 1.0%
Total Standard Uncertainty Components added in quadrature	± 1.7%	± 1.7%	± 1.6%	± 4.2%	± 1.7%
Expanded uncertainty *	± 3.4%	± 3.4%	± 3.2%	± 8.4%	± 3.4%

* The figures quoted for the uncertainty in the reference irradiation distance includes a sensitivity factor of 2, taking into account the inverse square dependence of the neutron fluence rate on the distance between the source centre to reference point.

⊗ Obtained by multiplying the total standard uncertainty by a coverage factor $k=2$. (This provides an uncertainty estimate at a confidence level of approximately 95%.)