Christian.Hranitzky@seibersdorf-laboratories.at Coordinator of IC2021area







Feedback and Conclusions

Intercomparison IC2021area of passive area dosimetry systems

Christian Hranitzky, Christian Gärtner, Christian Naber, Julia Aslan, Maria A. Duch, Thomas Haninger, Željka Knežević

Conclusions I




- 1 participant could not provide dosemeters to KIT in time
- 1 participant could not report dose results
- 7 systems needed a bag for outdoor hanging positions



water inside the bag


Preliminary average results – 65 systems






Preliminary average results – 28 systems: 3-months outdoor






Preliminary average results – 23 systems: 3-months indoor





Preliminary average results – 14 systems: 6-months outdoor





Conclusions II



Most participating dosimetry systems performed well:

93 % of the response results within the ISO limits Similar overall results for the 3 measurement conditions Similar average response for 150 µSv and 300 µSv

Radiation quality Cs-137 was used to focus on:

Low dose measurements: 150 μSv and 300 μSv Background dose values: 200 μSv - 1600 μSv

***** KIT calibration laboratory performed well:

1 irradiation was necessary to repeat Photo documentation of all irradiations was perfect

Response Limits Criteria – ISO 14146



Radiological protection — Criteria and performance limits for the periodic evaluation of dosimetry services

7.1.1 Personal and area dosemeters

For each irradiated dosemeter, the quotient R between the measured dose value G and the conventional quantity value H_{ref} , given by the response, as in Formula (1):

$$R = \frac{G}{H_{\text{ref}}} \tag{1}$$

shall meet the following criteria between H_0 and H_{top} (see <u>6.3</u>):

— Criterion 1) For photon radiation with a mean energy of $\overline{E}_{\rm ph} > 10$ keV and for beta radiation with a mean energy of $\overline{E}_{\rm beta} > 0.2$ MeV (easier-to-measure):

$$0,71 \cdot \left(1 - \frac{2 \cdot H_0 / 1,33}{H_0 / 1,33 + H_{\text{ref}}}\right) \le R \le 1,67 \cdot \left(1 + \frac{H_0}{4 \cdot H_0 + H_{\text{ref}}}\right) \tag{2}$$

Conclusions III



Current revision of ISO 14146 triggered by IC2021area:

appropriate response limits useful requirements e.g. for reference conditions

Necessary information about the type of application:

environmental monitoring workplace monitoring environmental & workplace monitoring

Online Platform

https://www.eurados-intercomparison.org





EURADOS Intercomparison Platform

Start page

Extremity & eye lens dosemeter intercomparison IC2019exteye

Whole body dosemeter intercomparison IC2020ph

Passive area dosemeter intercomparis on IC2021area

Neutron dosemeter intercomparison IC2022n


Whole body dosemeter intercomparison IC2022ph

Online Platform - Dosimetry System LABORATORIES



System Identification:	TLD-700 Please describe the Dosimetry System (e.g. model, reader system, accreditation number,). This text will be included in the certificate.	0
Measurement Quantity:	H*(10) Please specify the measurement quantity in which you will report the dose values of the irradiated dosemeters.	0
Detector Type / Materials:	TLD-700 (LiF: Mg, Ti) Please specify the detector type (e.g. TLD, OSL,) and materials (e.g. LiF:Mg,Ti, TLD700,) of your dosemeter system for statistical analysis.	
Number of Detectors:	4 Please specify the number of detector elements within your dosemeter for statistical analysis.	
Reference Energy / Calibration Radiation Quality:	Cs-137 Please specify the reference energy / calibration radiation quality of your dosemeter system.	
Measurement Condition:	outdoor, 6 months Please choose one of three measurement conditions for your dosemeter system to be used for the IC2021area intercomparison (indoor or free-field outdoor measurement position, 3 months or 6 months measurement period).	
Lower Measurement Limit:	36–70 µSv Please choose the lower measurement limit (alternatively the lower detection limit) of your dosemeter system for statistical analysis.	0









Passive area dosemeter intercomparison IC2021area

Start page > Documents > Feedback > Logout (christian.hranitzky@seibersdorf-laboratories.at) Status A Monitoring Service Contact Person Dosimetry Systems Change password Tools

♦ Online Survey ♦

Please take a few minutes and complete the survey below. Your feedback is appreciated and helps us to improve the post intercomparison.

18 Participants answered:

Participant ID:	+	X	-
01. Has the intercomparison organisation fulfilled your expectations?	17	1	0
02. Have your own intercomparison results fulfilled your expectations?	14	4	0



03. Any comments on the price of participation?	
04. Any comments on the usage of the online platform or the communication with the coordinator?	
05. Any comments on the 'Instructions for Participants' or 'Terms and Conditions'?	
06. Any comments on the time schedule (registration, sending/receiving of dosemeters, measurement periods, dose reporting, draft/final certificates, EURADOS report)?	
07. Any comments on the irradiated low dose values (compared to the natural background + transport dose contribution)?	
08. Any comments on the measurement conditions (indoor and outdoor, 3 and 6 months), number of dosemeters or dosemeter positioning?	•••
09. Any comments on the 'Certificates of Participation' or the EURADOS Report?	



03. Any comments on the price of participation?	4	0	0
04. Any comments on the usage of the online platform or the communication with the coordinator?	10	0	0
05. Any comments on the 'Instructions for Participants' or 'Terms and Conditions'?	6	0	1
06. Any comments on the time schedule (registration, sending/receiving of dosemeters, measurement periods, dose reporting, draft/final certificates, EURADOS report)?	8	0	0
07. Any comments on the irradiated low dose values (compared to the natural background + transport dose contribution)?	3	1	1
08. Any comments on the measurement conditions (indoor and outdoor, 3 and 6 months), number of dosemeters or dosemeter positioning?	4	1	0
09. Any comments on the 'Certificates of Participation' or the EURADOS Report?	7	0	0 "



"These doses are so low that it becomes difficult to determine precisely their value. This is why the whole exercise is interesting"

"The graph (figure 5.5) with the ranking number is a good idea"

"We had irradiation during the transit, so the range doses (150- $300\mu Sv$) were very low for us"

"It's not clear if a distinction is made to 'transport dose' and 'natural background dose' subtraction. Look at point 7. So if a service as a way of determining natural background it was subtracted from the results and only 'transport dose' was 'left' on the non-exposed dosemeters. But if transport dose was considered as the sum of both no correction at all was done"

Questionnaire 2016

EURADOS Report 2021-02







Figure 16: Dose calculation methods.

Background dose

SEIBERSDORF LABORATORIES


Average of 6 not-irradiated dosemeters



Background dose

SEIBERSDORF LABORATORIES

Average of 6 not-irradiated dosemeters



EURADOS Report 2022-01

http://eurados.org

https://eurados.sckcen.be



