

Intercomparison IC2021 area of passive area dosimetry systems – a review by organizers and participants (WG3)

Results and measurement uncertainty of the CIEMAT TLD system in the IC2021 area intercomparison

I. Introduction

II. CIEMAT TLD system for Workplace and Environmental Dosimetry

III. Dose evaluation and measurement uncertainty

IV. Results in the IC2021 area intercomparison

V. Conclusions

rafael.rodriguez@ciemat.es
Ionizing Radiation Dosimetry Unit
CIEMAT

Keywords: *Environmental Radiation Monitoring, passive dosimetry, area dosimeter, measurement uncertainty, Intercomparison*

I. Introduction


➤ Background: CIEMAT participation in the two previous ICs organized by WG3-SG2

□ Four references sites (PTB): Free Field, Platform, UDO-II and **Cs-137 source**

- IC2014env, irradiations:
 - ✓ Free field (3 or 6 months)
 - ✓ SCR at the lake platform
 - ✓ **Cs-137 (0°): 5,5mSv**
 - ✓ Transport dose: UDOII

- IC2017prep, irradiations:
 - ✓ Free field (6 months)
 - ✓ SCR at the lake platform
 - ✓ **Cs-137 (0°, 90°): 30mSv**
 - ✓ Transport dose: UDOII

Ciemat TLD results according to ISO-14146:2018



I. Introduction


➤ Last EURADOS Intercomparison organized by WG3-SG2: **IC2021area** (from Jul-21 to Feb-22):

- Outdoor location:
 - ✓ Outdoor 3 / 6 moths
 - ✓ Cs-137 irradiations (0°)
- Indoor location:
 - ✓ Inside 3 moths
 - ✓ Cs-137 irradiations (0°)


□ 2 irradiations at doses lower than 0.5mSv in the KIT calibration laboratory

- 6 dosimeters for exposure outdoor/indoor in free field
- 2 groups of 3 dosimeters each for low dose irradiation: 300 μ Sv and 150 μ Sv
- Reported doses by labs without transport dose correction

II. CIEMAT TLD system for Workplace and Environmental Dosimetry




Detector handling



PTW Oven

- Annealing
- Pre-Reading cycle


⁹⁰Sr-Y Irradiator



Harshaw 5500 Reader

CIEMAT Environmental Dosimetry Laboratory

- Outdoor for environmental monitoring
- Indoor for workplace monitoring




TLD materials: LiF:Mg,Cu,P (GR-200) and LiF:Mg,Ti (TLD-100)

II. CIEMAT TLD system

for Workplace and Environmental Dosimetry

Dosimetry system features

- Validation of influence factors based in the IEC 62387 standard
- Traceability and QC procedure:
 - ✓ $H^*(10)$, Cs-137 annual calibration in a metrology laboratory
 - ✓ QC process: ICFs calculation (after reading)
- Regular participation in the spanish IC (every 5 years)
- Accreditation according to ISO 17025, since April 2012



Ciemat
Centro de Investigaciones Energéticas, Medioambientales y Tecnológicas
Servicio Instituto de Radiaciones

ENAC
ENSA Y O S
Nº 144LE1836

Código Informe: PT-LDA-200-R36602021
Página 1 de 1

LABORATORIO DE DOSIMETRÍA AMBIENTAL Y DE AREA (LDA)
Av. Complutense nº 40. Edificio 36. 28040 Madrid

INFORME DE RESULTADOS
DOSIMETRÍA AMBIENTAL TERMOLUMINISCENTE

DATOS DEL CLIENTE: EURADOS Intercomparison IC2021area
Att. Christian Hranitzky
IC2021area Coordinator
e-mail: ic2021area@eurados-intercomparison.org

Método utilizado: PG LDA 001, PT LDA 200

Resultados de la evaluación de 12 dosímetros de área procedentes de la instalación:
EURADOS Passive Area Dosimeter Intercomparison IC2021area

F. Borrado: 07/07/2021
F. Recepción: 26/10/2021
F. Lectura: 29/10/2021

Estación	Chasis	Período Exposición	Dosis $H^*(10)$ (μSv)	Tasa dosis (mSv/h)	Incertidumbre U ($k=2$)	Dosis anual ^a (mSv)	NOTAS
1	C012/2021-01	07-jul-21 / 29-oct-21	415	152 ± 9			
2	C012/2021-02	07-jul-21 / 29-oct-21	258	94 ± 9			
3	C012/2021-03	07-jul-21 / 29-oct-21	228	83 ± 10		1,33	
4	C012/2021-04	07-jul-21 / 29-oct-21	561	205 ± 12		0,73	
5	C012/2021-05	07-jul-21 / 29-oct-21	243	89 ± 4		1,80	
6	C012/2021-06	07-jul-21 / 29-oct-21	559	204 ± 17		0,78	
7	C012/2021-07	07-jul-21 / 29-oct-21	251	92 ± 3		1,79	
8	C012/2021-08	07-jul-21 / 29-oct-21	398	146 ± 10		0,80	
9	C012/2021-09	07-jul-21 / 29-oct-21	257	94 ± 16		1,27	
10	C012/2021-10	07-jul-21 / 29-oct-21	569	206 ± 20		0,82	
11	C012/2021-11	07-jul-21 / 29-oct-21	412	150 ± 17		1,32	
12	C012/2021-12	07-jul-21 / 29-oct-21	240	88 ± 4		0,77	

$H^*(10)$: Dosis Equivalente Ambiental neta, calculada con deducción de dosis de tránsito
Valor empleado como dosis de tránsito: 0 μSv (No se aplica, por indicación del organizador)
^a Dosis Anual: Dosis acumulada en la hipótesis de permanencia continua en la estación durante 1 año natural completo (8760 h)

Madrid, 24 de noviembre de 2021

III. Dose evaluation and measurement uncertainty

- Two methods based on GUM (JCGM 100:2008), linked to ISO/IEC Guide 98-3 (Part 3)
 - a) A general method based on the laboratory methods
 - Influence factors on the expanded measurement uncertainty $U'(k=2)$:
 - ✓ Reader counts (2%)
 - ✓ Reader calibration factor (4%)
 - ✓ Energy response (2%) and angular response (5%)
 - ✓ Calculation of ICFs (3%), ^{90}Sr -Y Irradiator
 - ✓ Transport dose subtraction (depends on the trip)
 - Expanded measurement uncertainty: 16%, without trip dose subtraction
 - b) Method based on the standard deviation (n° detectors / dosimeter):


Dose mean value, standard deviation ($n-1$) and t-distribution

 - IC2021area: TLDs uncertainties ($k=2$), range [4, 13]%, without trip dose subtraction

III. Dose evaluation and measurement uncertainty

- Dosimeter reading (counts)
- Reader calibration factor (trazability)
- Detector calibration (ICFs calculation), $\mu\text{Sv}/\text{count}$
- Dose evaluation, $H^*(10)$

Net Doses (SGCA)



Calculation of the $H^*(10)$ doses

ICFs calculation				Corrected counts				Dose evaluation				Calibration factor (25/11/2020)			
FCI ($\mu\text{Sv}/\text{cta}$)	Estación	Dosimetro	Observ.	Lectura Neta (cta)	Dosis (μSv)	Media (μSv)	STD (μSv)	CV	U' (k=1) H*(10), μSv	Tasa Dosis (nSv/hora)	H*(10) (μSv)	CV U (k=1) (μSv)	Tasa Dos: U (k=2) (nSv/hora)	CV (nSv/hora)	
0,00052	1	1		789381	412				1,00						
0,00055	C012/2021-01	2		784017	428										
0,00053	15-jul-21	3		798030	421				2,8%	182	415	2,8%	182	6%	
0,00055	18-oct-21	4		731894	401	415	11,7	2,8%	415	182	415	2,8%	182	10%	
0,00102	2	5		264865	271				1,00						
0,00081	C012/2021-02	6		317571	258										
0,00094	15-jul-21	7		275605	260				4,9%	113	258	4,9%	113	11%	
0,00087	18-oct-21	8		276281	241	258	13	4,9%	258	113	258	4,9%	113	11%	
0,00062	3	9		354609	220				1,00						
0,00074	C012/2021-03	10		305059	226										
0,00068	15-jul-21	11		364412	247				5,7%	100	228	5,7%	100	11%	
0,00053	18-oct-21	12		412312	219	228	13	5,7%	228	100	228	5,7%	100	11%	
0,00052	4	13		1034557	542				1,00						
0,00050	C012/2021-04	14		1149441	580										
0,00056	15-jul-21	15		1001574	560				2,8%	246	561	2,8%	246	6%	
0,00056	18-oct-21	16		1004474	564	561	15,7	2,8%	561	246	561	2,8%	246	14%	

IV. Results

in the IC2021 area intercomparison

4. Table of results

In the upper part of the following table, the reported dose values M_0 of the 6 background dosimeters and the resulting average background dose \bar{M}_0 are listed. In the lower part, the reported dose values M of the 6 irradiated dosimeters as well as the background corrected dose values $(M - \bar{M}_0)$, the reference dose values H_{ref} and the resulting dosimeter response values $R = (M - \bar{M}_0) \cdot H_{ref}^{-1}$ are shown.

Dosimeter	M_0
ID	μSv
S012-03	228
S012-12	240
S012-05	243
S012-07	251
S012-09	257
S012-02	258
Average	246

Background dosimeters

Dosimeter results for low doses

Dosimeter	M	$M - \bar{M}_0$	H_{ref}	R
ID	μSv	μSv	μSv	1
S012-08	398	152	150	1,01
S012-11	412	166	150	1,11
S012-01	415	169	150	1,13
S012-06	559	313	300	1,04
S012-04	561	315	300	1,05
S012-10	569	323	300	1,08


Cs-137 response for 0,3mSv and 0,15mSv

IV. Results

in the IC2021 area intercomparison

Background (outdoors) of the 6 TLDs (transport dose integrated) for 3 months


- ✓ Dose values in the range [-7, +5]% referred to the mean value
- ✓ Uncertainties $U'(k=2)$ in the range 4% - 13%
- ✓ Mean dose value (6 dosimeters): 246 (47) μSv ; $U'(k=2) = 19\%$



IV. Results

in the IC2021 area intercomparison

Cs-137 irradiations of 3 dosimeters (background dose subtracted) for low doses < 0,5mSv



Referred to $H^*(10) = 150\mu\text{Sv}$:

- Dose values in the range [+1, +13]%
- Uncertainties $U'(k=1)$ in the range [11, 17]%


Referred to $H^*(10) = 300\mu\text{Sv}$:

- Dose values in the range [+4, +8]%
- Uncertainties $U'(k=1)$ in the range [8, 10]%

IV. Results

in the IC2021 area intercomparison

Cs-137 irradiations of 3 dosimeters (background dose subtracted) for low doses < 0,5mSv



Mean dose values:


- 0,162 (18) mSv, $U'(k=1) = 6\%$
- 0,32 (3) mSv, $U'(k=1) = 17\%$

Referred to $H^*(10) = 150\mu\text{Sv}$:


- Dose values in the range [+1, +13]%
- Uncertainties $U'(k=1)$ in the range [11, 17]%

Referred to $H^*(10) = 300\mu\text{Sv}$:

- Dose values in the range [+4, +8]%
- Uncertainties $U'(k=1)$ in the range [8, 10]%




$H_{ref} : U'(k=2) = 4,3\%$




IV. Results

in the IC2021area intercomparison

Calibration stability of the Reader (Harshaw 5500)



→ Stability of the calibration factor



← EURADOS ICs:
Linearity from 0,15mSv to 30mSv

CIEMAT TLD system in the three EURADOS ICs

V. Conclusions

- Cs-137 response [+5, +8]% in the three EURADOS ICs (2014-2021)
- Linear dose response (0,15 - 30 mSv) to Cs-137 has been checked
- Measurement uncertainty are consistent by the two methods