

## **Project EIVIC2020 (EC, DG-ENER)**

# **“European In-vivo Counting Intercomparison Exercise 2020-2021“**

**WP. 2.3 – ANALYSIS OF RESULTS  
T. BEAUMONT (IRSN)**

**EC Contract ENER/2019/NUCL/SI2.811157**

**WEBINAR EURADOS 04.12.2023**

# INTRODUCTION

## Objective of the WP. 2.3 – Collecting and analysis of results

- ✓ Developing a specific template dedicated to collect results
- ✓ Collecting and checking data of participants
- ✓ Performing statistic analysis of results according to the International Standard (ISO)
- ✓ The conformity and performance of the facilities for *in vivo* monitoring

## The EIVIC Intercomparison

- ✓ 41 facilities from 35 laboratories and 21 countries
- ✓ 5 configurations covering the range of such possible measurements associated to different intake scenarios
  - ✓ Task 1 – Victor (*suitable for NaI(Tl) and germanium detectors: P5 phantom*)
  - ✓ Task 2 – Emergency (*suitable for NaI(Tl) and germanium detectors: P4 phantom*)
  - ✓ Task 3 – Medicine (*suitable for NaI(Tl) and germanium detectors: P5 phantom*)
  - ✓ Task 4 – Calibration (*suitable only with germanium detectors: P4 and P5 phantom*)



## Date of the beginning of the analysis

- ✓ 5<sup>th</sup> May 2022 : receipt of the latest results

# PROCESS OF DATA EVALUATION (1/2)


- Downloading of the measurement results (type of radionuclides and activity) by the participants on the **BfS Cloud**.



- Use a link between the download templates on the cloud and the IRSN analysis template:
  - To limit the input error
  - To reduce the check

Laboratory	Ass. No.	S	F	N	Date of measurement	Measured activity (Bq)	Uncertainty (%)	LD (Bq)	Activity at reference date (Bq)	Measured activity (Bq)	Uncertainty (%)
28.V					1000	151	89	1082	2650	950	
29.V	640	300	84	975	2700	700					
3.6					1000	100	57	2039	2400	960	
20.S					1180	140	41.2	1296	2810	290	
21.V	1082	87	77	1138	2900	100					
28.V	1142	59	33	1178	2838	95					
27.V	1066	213	409	888	1470	211					
20.V	1018	100	95	1057	1417	170					
4.S	1258	228	88	1299	3204	289					
9.S	1002	309	58	1097	2468	170					
24.S	1056	105	68	1114	2653	939					
23.V	844			872	3772						
33.V	1138	78	61	1192	3020	216					
8.V	1252	124	586	1296							
28.V					1110	384	152	1383	2889	939	
24.S	893	268	78	962	2465	187					
31.S	1384	204	170	1480	4080	717					
17.S	1256	172	86	1296	2615	342					
28.V	491	244	100	533	1380	144					
13.V	1580	200	100	1654	3680	400					
18.V	1200	394	32.2	1301	2899	347					
7.V	1342	280	96	1391	2720	460					
39.V	2100	432	340	2234	6780	818					
3.6	1011	104	184	1077	5007	106					

- Statistical treatment by proficiency tests carried out with **Prolab™ Software** according to international standards **ISO 13528** and **ISO 28218**




# PROCESS OF DATA EVALUATION (2/2)

## Assigned value (ISO 13 528):

There are many options to determine the assigned values, as example used of:

- ✓ The reference value of the certificate
- ✓ The consensus value from expert laboratories (arithmetic mean)



- Only 3 experts laboratories : KIT, CIEMAT and IRSN
- **Task 1, 2 and 4** : measurements performed by 3 experts
- **Task 3**: nuclear medicine (*short half-life radionuclide*), measurement performed by **1 expert** (CIEMAT)

- ✓ The consensus value from participant results (determined with robust method)

Lot of data : 41 facilities

The robust mean was used to guarantee a **homogeneous analyze** between tasks.

Task 2– Emergency	<sup>137</sup> Cs
Reference value (BFS)	3225 ± 42
Robust mean	2996.18 ± 60.40
Difference (%)	-7.09%


# STATISTICAL TREATMENT

## Bias (ISO 28 218) :

According to the recommendations of **ISO 28218** "Performance criteria for radiobioassay", the **relative bias** statistic shall be **between [-25% to +50%]** relative to the target value. The relative bias statistic is defined as:

$$Bias(\%) = \frac{a - A}{A} \times 100$$

*a*: Result of the participating facility  
*A*: Activity of the assigned value




## Statistic test on the distribution of results: the research of outliers (Grubbs test)

## Z-score (ISO 13 528):

The Z-score is **an indicator of the laboratory proficiency** compared to that of the other laboratories. According to the recommendations of ISO 13528 "Statistical methods for use in proficiency testing by intercomparison", the current Z-score criteria are:

- ✓  $|Zscore| \leq 2$ : the result is acceptable;
- ✓  $2 < |Zscore| \leq 3$ : The result give a warning signal;
- ✓  $|Zscore| > 3$ : The result give an action signal.



The Z-score depends **directly on the dispersion of the results from the laboratories.**

# EXAMPLES OF RESULTS

TASK 1 - VICTOR

TASK 2 - EMERGENCY

TASK 3 - MEDICINE 1&2

TASK 4A - CALIB.

TASK 4B - CALIB.

CONCLUSION

## RESULTS

# TASK 2 : EMERGENCY

$^{134}\text{Cs}$ ,  $^{137}\text{Cs}$ ,  $^{40}\text{K}$   
P5 90KG


# RESULTS – RAW DATA OF PARTICIPANTS

## Assigned value (ISO 13 528):

[ <sup>137</sup>Cs ]

Number of facilities reporting results: 40

Statistic robust method: Q/Hampel



Task 2– Emergency	<sup>137</sup> Cs
Reference value (BFS)	3225 ± 42
Robust mean	2996.18 ± 60.40
Difference (%)	-7.09%


# RESULTS - BIAS

## Bias (ISO 28 218) :

[ <sup>137</sup>Cs ]

Task: : Emergency phantom 90 kg  
Radionuclide: cesium 137

Assigned value: 2996,181 Bq (empiric value)



PROLabPlus




Id	<sup>137</sup> Cs
28	-51.87%
30	-41.16%
39	114.04%

**3 facilities [28, 30 and 39] are in nonconformity**  
**37 facilities are conform**  
 according to the ISO 28 218 criteria '[-25%:+50%]'



# RESULTS – Z-SCORE



$|Zscore| \leq 2$ : acceptable

$2 \leq |Zscore| \leq 3$ : warning signal

$|Zscore| > 3$ : action signal

The z-score was calculated and according to the Z-score criteria of ISO 13528, there are:

- ✓  $^{134}\text{Cs}$ :
  - ✓ 34 facilities 'acceptable'
  - ✓ 6 'unacceptable'
- ✓  $^{137}\text{Cs}$ :
  - ✓ 35 facilities 'acceptable',
  - 1 'warning signal'
  - ✓ 4 'unacceptable'
- ✓  $^{40}\text{K}$ :
  - ✓ 25 facilities 'acceptable',
  - 1 'warning signal'
  - ✓ 2 'unacceptable'

# CONCLUSION ON THE RESULTS FOR THE TASK 2

Id	Task 2 - Emergency	
	ISO 28 218	ISO 13 528
1	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
2	Conform	Acceptable
3	Conform	Acceptable
4	Conform	Acceptable
5	Conform	Acceptable
6	Conform	Acceptable
7	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
8	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
9	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
10	Conform	Acceptable
11	Conform	Acceptable
12	Conform	Acceptable
13	Conform	Acceptable
14	Conform	Acceptable
15	Conform	Acceptable
16	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
17	Conform	Acceptable
18	Conform	Acceptable
19	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
20	Conform	Acceptable
21	Conform	Acceptable
22	Conform	Acceptable
23	-	-
24	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
25	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)

Id	Task 2 - Emergency	
	ISO 28 218	ISO 13 528
26	-	-
27	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>137</sup> Cs) Action signal ( <sup>134</sup> Cs) Warning signal ( <sup>40</sup> K)
28	Not conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs/ <sup>40</sup> K)	Action signal ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
29	Conform	Acceptable
30	Not conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Action signal ( <sup>134</sup> Cs/ <sup>137</sup> Cs) Warning signal ( <sup>137</sup> Cs)
31	Conform	Action signal ( <sup>134</sup> Cs)
32	Conform	Acceptable
33	Conform	Acceptable ( <sup>40</sup> K) Action signal ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
34	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
35	Conform	Acceptable
36	Conform	Acceptable
37	Conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs)	Acceptable ( <sup>134</sup> Cs/ <sup>137</sup> Cs)
38	Not conform ( <sup>40</sup> K)	Action signal ( <sup>40</sup> K)
39	Conform	Acceptable
40	Not conform ( <sup>134</sup> Cs/ <sup>137</sup> Cs/ <sup>40</sup> K)	Action signal ( <sup>134</sup> Cs/ <sup>137</sup> Cs/ <sup>40</sup> K)
41	-	-
42	Conform	Acceptable
43	Conform	Acceptable

The tolerance intervals are **more restrictive** according to ISO 13528 (bias) than to ISO 28218 (Z-score)

# STATISTICAL TESTS (1/3)

---

## STATISTICAL TESTS REGARDING INFLUENCING PARAMETERS

### Measurements not carried out under equal conditions and with equal installations :

- Detection system : NaI(Tl) or HPGe detector
- Type of participation: attended tour or shipment
- Type of calibration phantom : more or less realistic phantom
- Calibration curves (70 kg systematic or adapted)
- Measurement geometry: sitting, lying or standing.
- Duration of the measurement
- Detector-subject distances

### Statistical tests were performed using R software and regarding:

- ✓ The central tendency:  
**Mann-Whitney U test** “Do values A tend to be greater or smaller than values B?”
- ✓ The dispersion:  
**Siegel-Tukey test** “Do values A tend to be more or less dispersed than values B?”
- ✓ **Data used**: all reported z-scores (except outliers) from all **tasks (1-4)** that involved phantoms

# STATISTICAL TESTS (2/3)

## MEASUREMENT OF A 90 KG PHANTOM WITH A 70 KG CALIBRATION CURVE


Task 2 'Emergency' used a phantom of **90 kg**

### Two calibration methods:

- 70 kg calibration curve applied systematically for measuring all sizes of people and phantoms
- Adapted calibration curves applied for measuring

### Results:

- Significantly different when the 70kg calibration curve is systematically used. ( $p=2,5E-4$ )



Z-score discriminated according to the phantom masses for calibration measurements  
(data of Task 2 'Emergency')

# STATISTICAL TESTS (3/3)

## TYPE OF CALIBRATION PHANTOM


Bottle phantom vs. brick phantom:

- Similar results with the central tendency (p-value 0,59)
- Different results with the dispersion (p-value 9,3E-4)

Results:

- with the **Canberra phantom** tended to be **underestimated** (p-value 3,5E-4 )
- with **own phantoms** tended to be **overestimated** (p-value 6,3E-4 )

→ Small dispersion : different makeups of these phantoms



Z-score discriminated according to the type of calibration phantom  
(all data)

# ERRATUM : TASK 3 MEDICINE #1 & #2

TASK 1 - VICTOR    TASK 2 - EMERGENCY    **TASK 3 - MEDICINE 1&2**    TASK 4A - CALIB.    TASK 4B - CALIB.    CONCLUSION

## RESULTS

### TASK 3 : MEDICINE #1

$^{68}\text{Ge}/^{68}\text{Ga}$ ,  $^{88}\text{Y}$ ,  $^{40}\text{K}$   
P4 70KG

FACILITY NUMBER : 1, 3, 4, 5, 6, 8, 10, 13, 17, 22, 25, 30, 32, 34, 37, 38, 41, 42, 43

**4 facilities [3, 6, 10 and 17]  
move from Task 3.1  
to Task 3.2**

TASK 1 - VICTOR    TASK 2 - EMERGENCY    **TASK 3 - MEDICINE 1&2**    TASK 4A - CALIB.    TASK 4B - CALIB.    CONCLUSION

## RESULTS

### TASK 3 : MEDICINE #2

$^{68}\text{Ge}/^{68}\text{Ga}$ ,  $^{88}\text{Y}$ ,  $^{40}\text{K}$   
P4 70KG

FACILITY NUMBER : 2, 7, 9, 11, 12, 14, 18, 19, 20, 21, 24, 27, 28, 29, 31, 33, 35, 36 AND 39

# ERRATUM : TASK 3 MEDICINE #1 & #2

## Results

Task 3.1 – Medicine		Ga-68	Y-88	K40
NEW ANALYSIS	Assigned Value (Bq)	3741	4707	3862
	Uncertainty (Bq)	114	163	127
PREVIOUS ANALYSIS	Assigned Value (Bq)	3741	4852	3812
	Uncertainty (Bq)	114	194	127
$\delta$   (%) assigned value		-	<b>3.1</b>	<b>1.3</b>

- Differences are acceptable : 3.1% and 1.3%.
- Bias and z-score values slightly changed for all facilities expect for:
  - idLab 32 (<sup>40</sup>K): acceptable (z-score = -1.8) and now, warning signal (z-score = -2.2).

Task 3.2 – Medicine		Ga-68	Y-88	K40
NEW ANALYSIS	Assigned Value (Bq)	4219	4263	4163
	Uncertainty (Bq)	151	127	201
PREVIOUS ANALYSIS	Assigned Value (Bq)	4219	4283	4313
	Uncertainty (Bq)	151	171	210
$\delta$   (%) assigned value		-	<b>0.5</b>	<b>3.6</b>

- Differences are acceptable : 0.5% and 3.6%.
- **4 facilities** [3 (<sup>88</sup>Y, <sup>40</sup>K); 6 (<sup>88</sup>Y, <sup>40</sup>K), 10 (<sup>40</sup>K) and 17 (<sup>88</sup>Y, <sup>40</sup>K)] are **now conform** (bias) and **acceptable** (z-score).
- Bias and z-score values slightly changed for all facilities expect for:
  - idLab28 (<sup>88</sup>Y) : warning signal (2.7) and now, action signal (z-score=3.5)
  - idLab14 (<sup>40</sup>K) : warning signal (-2.1) and now it is acceptable (z-score=-1.8)
  - idLab39 (<sup>40</sup>K) : conform (+47.2%) and now it is not conform (bias = +52.5%)

# ERRATUM : TASK 3 MEDICINE #1 & #2

- **Publications and communications on the Project take account these modifications:**

- Publishable Reports (EURADOS and RP Report)

<https://eurados.sckcen.be/news-overview/eurados-report-2023-03-published-eivc-2020-european-vivo-intercomparison-exercise-2020-organisation-european-interlaboratory-comparison-whole-body-counting>

- Journal(s)/conference(s):

- Franck et al, Radiat Environ Biophys submitted,
- Beaumont et al., IRPA (2024)



- **Erratum of participation certificates for Task 3 will be sent next week**





# GENERAL CONCLUSION

---

- ✓ **Objective:** to assess the implementation of the individual monitoring requirements in EU based on *in-vivo* measurements and receive an overview of the capabilities and performance of whole-body counters in Europe
- ✓ Measurements were carried out for **43 facilities** from 35 laboratories (+ European Commission and IAEA) and 21 countries (Only 2 facilities have not returned their results)
- ✓ Representing a very **important data base of European Laboratories** ( 5 exercises in 1 IC campaign)
- ✓ In general, there is a high participation:
  - ✓ **Task 1:** 40/41 facilities ( $^{137}\text{Cs}$ )
  - ✓ **Task 2 :** 40/41 facilities ( $^{134}\text{Cs}$ )
  - ✓ **Task 3:** 17/41 facilities ( $^{68}\text{Ga}$ ) more difficult because of not classical radionuclides.
  - ✓ **Task 4a:** 20/41 facilities ( $^{152}\text{Eu}$ ), explain because this task was dedicated to germanium detector and not mandatory
  - ✓ **Task 4b:** 30/41 facilities ( $^{152}\text{Eu}$ ), explain because this task was dedicated to germanium detector

# GENERAL CONCLUSION

---

- ✓ The **analysis of results** was carried out and the **conformity report of the facilities** are given, for each configuration, according to the **criteria** of **ISO 28218** and **ISO 13528**.
- ✓ **Results are quite good** (bias and z-score evaluation) and the most part of the facilities are conform according to the standards
- ✓ The EIVIC team could discuss with the laboratories in nonconformity to identify the sources of error
- ✓ Statistical tests were performed to test if they had a significant influence of the measurement parameters:
  - ✓ Surprisingly the results are quite similar for all the laboratories,
  - ✓ except for the phantom-size dependency, size-dependent calibration factors should be used.
- ✓ Erratum of participation certificates for Task 3 will be sent next week.

---

*Thank you for your attention*